

4

3

2

1

REVISION HISTORY				
ZONE	REV	DESCRIPTION	DATE	APPROVED

D

D

C


C

B

B

A

A



NOTES:

1. The chamber is 6'8" (203cm) wide by 6'6" (198cm) deep by 8'-0" (244cm) high.
2. The chamber weighs 1900 lbs (864 kg)
3. The Model 2100D is normally supplied with heavy duty casters. It can also be supplied on I-beam supports to spread the load over a wider area.
4. The Model 2100D requires an 120/240VAC 20/10A circuit for the chamber lights and accessories.
5. The instrumentation rack will require a separate circuit.
6. The chamber must be connected to electrical ground. A grounding stud is provided on the rear wall of the chamber.
7. Recommended minimum floor space is 12'6" (381cm) by 13'0" (396cm)

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES		THE HOWLAND COMPANY, INC.	
TOLERANCES:		4540 Atwater Court, Suite 107 Buford, GA 30518	
DECIMAL ANGULAR		678-546-5680 www.thehowlandcompany.com	
.X ±.050	±0.5°	TITLE MODEL 2100D WIRELESS TEST LAB	
.XX ±.030		FACILITY INTERFACE CONTROL DRAWING	
.XXX ±.010			
THIRD ANGLE PROJECTION		APVD	DATE
		DWN C SIRLES	DATE 11-25-14
		DSGN J HUFF	DATE 11-15-14
SIZE	CAGE CODE	DWG NO	REV
C	5V978	P0020003	
SCALE 1:15	SHEET 1 OF 1		